



MALTA RESOURCES AUTHORITY

MRA/108/2012

11th July 2011

TENDER FOR THE SUPPLY OF GROUNDWATER LEVEL MONITORING PROBES/EQUIPMENT FOR THE MORISO PROJECT

Question 1: In the tender lot 2 it is mentioned, that the depth range to groundwater is 200 m. We need to know the water level fluctuations, which is important to specify the measuring range of the sensor.

It is clarified that the depth range of 0 to 200m refers to the range of depths from the ground level to the water table in the identified groundwater monitoring boreholes in Malta. The fluctuations in the actual water-level is quite limited in range (0-10m).